

Study Spotlight: Rotational Stability ZEISS AT TORBI 709



Seeing beyond

A comparative peer-reviewed publication on rotational stability of two different IOL platforms in myopic eyes

Source



Title

Comparison of the rotational stability between plate-haptic toric and C-loop haptic toric IOLs in myopic eyes



Authors

Xiangjia Zhu, Jiaqi Meng, Wenwen He, Xianfang Rong, Yi Lu



Publication


Journal of Cataract and Refractive Surgery; October 2020; 46:1353–1359; doi: 10.1097/j.jcrs.0000000000000259

Methodology

Rotational Stability

- ZEISS AT TORBI 709 (plate-haptic toric IOL) compared to AcrySof Toric (C-loop haptic IOL) from Alcon in myopic cataract eyes
- Evaluating rotational stability, residual astigmatism, visual acuity, and higher-order aberrations (HOAs) at 3 months postoperatively

Sample Size



Two equally sized groups of 31 eyes

62 patients (62 eyes)

Results

- Plate-haptic toric AT TORBI® IOLs from ZEISS were more rotationally stable than the C-loop haptic toric AcrySof IOLs
- Less rotation corresponds with consequently reduced postoperative residual astigmatism and higher uncorrected visual acuity
- Plate-haptic toric IOLs might therefore be a better choice for myopic astigmatic cataract eyes because of less rotation